

UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Addiese: COMMISSIONER FOR PATENTS P O Box 1450 Alexandra, Virginia 22313-1450 www.wepto.gov

| APPLICATION NO. | FILING DATE | FIRST NAMED INVENTOR | ATTORNEY DOCKET NO. | CONFIRMATION NO. | |
|---|-------------|----------------------|---------------------------------|------------------|--|
| 10/615,636 | 07/10/2003 | Jay A. Warren | 279.044US1 | 3513 | |
| 21186 7550 07/09/2008 SCHWEGMAN, LUNDBERG & WOESSNER, P.A. | | | EXAM | EXAMINER | |
| P.O. BOX 2938 | | | ALTER MORSCHAUSER, ALYSSA MARGO | | |
| MINNEAPOLIS, MN 55402 | | ART UNIT | PAPER NUMBER | | |
| | | | 3762 | | |
| | | | | | |
| | | | MAIL DATE | DELIVERY MODE | |
| | | | 07/09/2008 | PAPER | |

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Application No. Applicant(s) 10/615.636 WARREN, JAY A. Office Action Summary Examiner Art Unit ALYSSA M. ALTER 3762 -- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --Period for Reply A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS. WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION. Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication. If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication - Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b). Status 1) Responsive to communication(s) filed on 24 January 2008. 2a) This action is FINAL. 2b) This action is non-final. 3) Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under Ex parte Quayle, 1935 C.D. 11, 453 O.G. 213. Disposition of Claims 4) Claim(s) 1-38 is/are pending in the application. 4a) Of the above claim(s) 24-38 is/are withdrawn from consideration. 5) Claim(s) _____ is/are allowed. 6) Claim(s) 1-23 is/are rejected. 7) Claim(s) _____ is/are objected to. 8) Claim(s) _____ are subject to restriction and/or election requirement. Application Papers 9) The specification is objected to by the Examiner. 10) ☐ The drawing(s) filed on 10 April 2003 is/are: a) ☐ accepted or b) ☐ objected to by the Examiner. Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a). Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d). 11) The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152. Priority under 35 U.S.C. § 119 12) Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f). a) All b) Some * c) None of: Certified copies of the priority documents have been received. 2. Certified copies of the priority documents have been received in Application No. Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)). * See the attached detailed Office action for a list of the certified copies not received.

PTOL-326 (Rev. 08-06)

1) Notice of References Cited (PTO-892)

Notice of Draftsperson's Patent Drawing Review (PTO-948)

Imformation Disclosure Statement(s) (PTC/S5/08)
Paper No(s)/Mail Date ______.

Attachment(s)

Interview Summary (PTO-413)
Paper No(s)/Mail Date.

6) Other:

Notice of Informal Patent Application

Art Unit: 3762

DETAILED ACTION

Response to Arguments

Applicant's arguments with respect to claims 1-23 have been considered but are moot in view of the new ground(s) of rejection.

Claim Rejections - 35 USC § 112

The following is a quotation of the second paragraph of 35 U.S.C. 112:

The specification shall conclude with one or more claims particularly pointing out and distinctly claiming the subject matter which the applicant regards as his invention.

1. Claim 7 stands as rejected under 35 U.S.C. 112, second paragraph, as being indefinite for failing to particularly point out and distinctly claim the subject matter which applicant regards as the invention. Claim 3 recites the "QRS complex", however claim 7 recites "a T-wave". It is unclear if the Applicant has additional sensing circuitry to detect the T wave or if the T wave is sensed with the same circuitry that detects the "QRS complex".

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless -

- (b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.
- Claims 1-7 and 11-23 are rejected under 35 U.S.C. 102(b) as being anticipated by Stotts et al. (US 5,161,529). Stotts et al. discloses a pacemaker that "allows

Art Unit: 3762

switching of the amplifier's bandpass frequency characteristic thereby selectively varying the cardiac signal frequencies subject to sensing"(abstract).

"The sense amplifier comprises a switched capacitor amplifier which allows the amplifier's bandpass frequency to be switched at will, thereby selectively varying the cardiac signal frequencies subjected to sensing. The effect is to provide a sense amplifier with a programmable frequency response" (col. 3, lines 8-11).

"Prior to the application of a pacing stimulus to the heart, the frequency response (bandpass characteristic) of the sense amplifier is selectively set to detect cardiac activity in the form of the intrinsic QRS pattern, in the same manner as is customary with pacemakers of the prior art. However, at the moment that a stimulating pulse is delivered to the heart, the sense amplifier is switched to a lower frequency bandpass to render it more responsive to an evoked potential (for example, the T wave). After a time interval during which an evoked response to the stimulus should have been detected (the presence or absence of such a response being indicative of capture or the lack of capture, respectively), the amplifier is switched back to the original bandpass characteristic to sense the intrinsic response"(col. 3, lines 14-28).

Furthermore, "the complete sense amplifier is a cascade of highpass and lowpass stages which produces a bandpass characteristic which has frequency programmability together with switching the sense amplifier to a higher gain setting, which is suitable to sense potentials evoked by the stimulus" (col. 4, lines 60-65).

Art Unit: 3762

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negatived by the manner in which the invention was made.

1. Claims 8-10 are rejected under 35 U.S.C. 103(a) as being unpatentable over Stotts et al. (US 5,161,529). Stotts et al. discloses the claimed invention except for the range of time for the first time period. It would have been obvious to one having ordinary skill in the art at the time the invention was made to modify the range of time for the first time period, since it has been held that where the general conditions of a claim are disclosed in the prior art, discovering the optimum or workable ranges involves only routine skill in the art. *In re Aller*, 105 USPQ 233 (see MPEP 2144.05). Furthermore, a modification to the range of time of the first time period would provide the predictable results of modifying the treatment to meet specific patient needs.

Conclusion

Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within

Art Unit: 3762

TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to ALYSSA M. ALTER whose telephone number is (571)272-4939. The examiner can normally be reached on M-F 9am to 4pm.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Angela Sykes can be reached on (571) 272-4955. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/George R Evanisko/ Primary Examiner, Art Unit 3762 /Alyssa M Alter/ Examiner Art Unit 3762